



Republic of the Philippines
Department of Environment and Natural Resources
ENVIRONMENTAL MANAGEMENT BUREAU
DENR Compound, Visayas Avenue, Diliman, Quezon City, 1116
Telephone Nos. 927-15-17, 928-20-96
Email: emb@emb.gov.ph
Visit us at <http://www.emb.gov.ph>

SUPPLEMENTAL/BID BULLETIN NO. 1

Supply, Delivery and Installation of Two (2) Units Autoclave

This clarification/addendum No.6903518-1 is issued to clarify/amend item/s in the issued ITB No. EMB-GOODS-NCB 2020-006. This Bid Bulletin shall form an integral part of the bid documents

Name of Project : **Supply, Delivery and Installation of Two (2) Units Autoclave**

Date : March 16, 2020

Announcing the clarification/amendments/correction on the issued bidding documents:

Old Technical Specifications	New Technical Specifications
AUTOCLAVE: <ul style="list-style-type: none">o Efficient 15-L steam condenser eliminates steam exhaust into the labo Maximum sterile temperature: 135 °C	AUTOCLAVE: <ul style="list-style-type: none">- Removed- Maximum sterile temperature: 135 ± 1°C

For guidance and information of all concerned.


MS. CONSOLACION P. CRISOSTOMO
BAC Chairperson

Received by the Bidder:

Date: _____