



Republic of the Philippines  
**Department of Environment and Natural Resources**  
ENVIRONMENTAL MANAGEMENT BUREAU  
DENR Compound, Visayas Avenue, Diliman, Quezon City, 1116  
Telephone Nos. 927-15-17, 928-20-96  
Email: [emb@emb.gov.ph](mailto:emb@emb.gov.ph)  
Visit us at <http://www.emb.gov.ph>

### SUPPLEMENTAL/BID BULLETIN NO. 1

#### **Supply, Delivery, Installation and Commissioning of 1 Set Ammonia Ion Selective Electrode; 2 Sets Chloride Meter; 2 Units Digital Bottle-Top Burette; 1 Set Nitrate Ion Selective Electrode and 1 Unit Ultrasonic Bath, Heated**

This Supplemental/Bid Bulletin No. 7186712-1 is being issued to clarify/amend item/s in the issued **ITB No. EMB-GOODS-NCB-2020-0012**. This shall form an integral part of the Bid Documents.

Name of Project : **Supply, Delivery, Installation and Commissioning of 1 Set Ammonia Ion Selective Electrode; 2 Sets Chloride Meter; 2 Units Digital Bottle-Top Burette; 1 Set Nitrate Ion Selective Electrode and 1 Unit Ultrasonic Bath, Heated**

Date : September 3, 2020

Announcing the clarification/amendment/correction on the issued bidding documents:

<b>Original Technical Specifications</b>	<b>Revised Technical Specifications</b>
CHLORIDE METER: ISE Specifications Range: 0 to 19999  Includes: Chloride combination selective electrode, reference filling solution, Ionic Strength Adjuster, Chloride calibration standard (0.1m NaCl, 100ppm Cl, 1000ppm Cl), electrode stand, universal power adapter, user manual, computer interface cable	CHLORIDE METER: ISE Specifications Range: 0.0001 to 19000 with selectable units of ppm, Molar, mg/L, %, ppb or none  Includes: Chloride combination selective electrode, reference filling solution, Ionic Strength Adjuster, Chloride calibration standard (0.1 M NaCl, 100ppm Cl, 1000ppm Cl, at least 400mL each), electrode stand, universal power adapter, user manual, computer interface cable
<b>Original Schedule</b>	<b>Revised Schedule</b>
Deadline for Submission of Bids: September 07, 2020, 9:00 a.m.	Deadline for Submission of Bids: September 14, 2020, 9:00 a.m. Record Section, AFMD Office, AFMD Building, DENR Compound
Opening of Bids: September 07, 2020, 10:00 a.m.	Opening of Bids: September 14, 2020, 10:00 a.m. Record Section, AFMD Office, AFMD Building, DENR Compound

For guidance and information of all concerned.

**ALBERT A. MAGALANG**  
BAC Chairperson

Received by the Bidder:

Date: \_\_\_\_\_