



SUPPLEMENTAL/BID BULLETIN NO. 1

Supply, Delivery, Installation and Commissioning of UV-Vis Spectrophotometer

This clarification/addendum No. **7221847-1** is issued to clarify/amend item/s in the issued **ITB No. EMB-GOODS-NCB-2020-0031** and shall form an integral part of the bid documents.

Name of Project : ***Supply, Delivery, Installation and Commissioning of UV-Vis Spectrophotometer***

Date : October 14, 2020

Announcing the clarification/amendment/correction on the issued bidding documents:

Old Technical Specifications	New Technical Specifications
<ul style="list-style-type: none">• Optical System: Dual beam	<ul style="list-style-type: none">• Optical system: Dual beam or any technology that is equivalent or better to dual beam
<ul style="list-style-type: none">• Stray light: >2 A or better	<ul style="list-style-type: none">• Stray light: >2 A or ≤0.1%T at 340nm; ≤1%T at 198nm
<ul style="list-style-type: none">• Keypads	<ul style="list-style-type: none">• Keypads or touchscreen
<ul style="list-style-type: none">• Includes: Auto cell changer suitable for cuvettes of different path length (1, 2, 5, 10 cm)	<ul style="list-style-type: none">• Includes: Holder/s for cuvettes of different path length (1, 2, 5, 10 cm) and pair of cuvettes of different sizes (1, 2, 5, 10 cm)
<ul style="list-style-type: none">• Comes with software application	<ul style="list-style-type: none">• Stand-alone and with software application

For guidance and information of all concerned.

ALBERT A. MAGALANG
Chairperson
EMB Bids and Awards Committee

Received by the Bidder:

Date: _____