



SUPPLEMENTAL/BID BULLETIN NO. 1

This clarification/addendum No. **9057305** is issued to clarify/amend item/s in the issued **ITB No. EMB-GOODS-NCB-2022-0010** and shall form an integral part of the Bid Documents.


Name of Project : ***One (1) Lot Supply and Delivery of Five (5) Units Multi-Parameter Water Quality Monitoring System (Water Quality Checker) with Consumables***

Date : October 11, 2022

Announcing the clarification/amendment/correction on the issued bidding documents:

| Old Requirements | New Requirements |
|---|---|
| Delivery Period 30-45 calendar days | 60 calendar days |
| Availability of Equipment: Readily available in the Philippines | Remove |
| Min Dissolved Oxygen 0-50ppm | 0-20ppm |
| Dissolved Oxygen Accuracy - ± 0.2 ppm | ± 0.2 ppm or 0-200% or $\pm 1\%$ of reading |
| Min. Temp - ($^{\circ}$ C) -5 | 0 $^{\circ}$ C |
| Min mV (mV) - 2000 | ± 1000 |
| Min TDS 0 to 100pt | 0 to 100 g/L |
| Buffer Recognition: DIN, NIST and USA | Remove |
| Data Logging (Points) 10,000 data sets | Minimum of 1,000 full sets including GLP to Maximum of 10,000 data sets |
| Turbidity Accuracy $\pm 3\%$ of reading | Minimum of $\pm 3\%$ of reading to maximum of $\pm 5\%$ of auto range |
| Consumables Ph4 Calibration Solution | Ph4, Ph7 and Ph10 |

For guidance and information of all concerned.


ENGR. ESPERANZA A. SAJUL
 Chairperson
 EMB Bids and Awards Committee

