



**Department of Environment and Natural Resources  
ENVIRONMENTAL MANAGEMENT BUREAU**

DENR Compound, Visayas Avenue, Diliman, Quezon City 1116  
Tel.Nos. (632)927-1517; 928-3725; Fax No. (632)920-2258  
Website: [www.emb.gov.ph](http://www.emb.gov.ph) Email: [recordsco@emb.gov.ph](mailto:recordsco@emb.gov.ph)



Management System  
ISO 9001:2015  
ISO 14001:2015



www.tuv.com  
ID 9105082327

**SUPPLEMENTAL/BID BULLETIN NO. 1  
ADDENDUM/CLARIFICATION**

This Supplemental Bid Bulletin is issued announcing clarifications/corrections and amendments on the issued/posted bidding document for the Supply, Delivery, Installation and Commissioning of Various Laboratory Equipment under EMB-ERLSD-ELSS-BLU, ITB No. EMB-GOODS-NCB-2023-0001 posted on PhilGEPS on November 5, 2022, with reference No. 9189548. This shall form and integral part of the bid document.

**Name of Project:** Supply, Delivery, Installation and Commissioning of Various Laboratory Equipment under EMB-ERLSD-ELSS-BLU, ITB No. EMB-GOODS-NCB-2023-0001

**Date** : November 15, 2022

Amendments/correction on the issued bidding document as follows:

<b>Bidding Document</b>	<b>Original Specification</b>	<b>Revised/Amended Requirements</b>
<b>AUTOCLAVE</b>  <b>Technical Specifications</b> (page 32-33)	- 50 Liter Mew Hiclave for Rapid Heating  Agar Dissolution Protocol: Initially at 60oC – 100oC for 1-60 min, then 45oC-60oC for up to 24 hours	50 Liter Autoclave for Rapid heating  Agar Dissolution Protocol: Initially at 40oC – 100oC for 1-60 min, then 45oC-60oC for up to 24-48 hours
<b>WATER BATH</b>  <b>Technical Specifications</b> (page 34)	2-stage safety through relay switch-off close to the set point (approx. +10oC) and mechanical temperature limiter TB at 135oC	2-stage safety through relay switch-off close to the set point (approx. +10oC) and mechanical temperature limiter TB at 100-135oC

For guidance and information of all concerned.

  
**ENGR. ESPERANZA A. SAJUL**  
BAC Chairperson

hr